

Ref. No.: MAKAUT-WB/2021-22/LAB-INST.-PHARMACEUTICAL/002/CORRIGENDUM

Dated: 29-11-2021

This is for information of all the bidders that following amendments /Corrigendum are being made in tender documents (Ref. No.: MAKAUT-WB/2021-22/LAB-INST.-PHARMACEUTICAL/002 dated 12-11- 2021). Item Name: Instruments. The Specification of the said item should be read as follows:

The last date for submission for the tender is extended till 05-12-2021.

Aerators	For Organ baths of 230V. 50 Hz. 2 speed control, fitted with air flow control valve and tubing.	05
Convulsimeter	Electroconvulsimeter. Provide 50 Hz. stimulus current variable from 0.25 mA to 360 mA. A solid state timer controlling the stimulus duration from 0.1 to 1 sec. in steps of 0.1 sec. A stimulus pilot indicator when the stimulus is applied. Three pairs of corneal electrodes of 4,6 & 8 mm. & a pair of eye clips	01
Digital Plethysmograph	DIGITAL PLETHYSMOGRAPH with LCD Display. for Direct Volume Displacement readout permitting simple off line functioning. with inbuilt computer based graphical and numerical mode. Upto 10 experimental data storage per experiment. Supplied with software having facility to save experimental subject data. Unique normal tap water based transducer. No special fluid or conductivity control needed. Simple user friendly operations.	01
Histamine Chamber	Compact Histamine Aerosol apparatus- clear perspex exposition chamber divided in two sections, glass nebulizer and dial type sphygmomanometer	01
Dissection tray and board	Frog Dissection Board- 23 x 15 cm Lead lined with fixing elastic bands Wax trays	03
Stereotaxic Apparatus Rat & Guinea Pig	100 micron 3-axes, left-hand manipulator arm, rat adaptor, traditional (18°) ear bars, and standard probe holder.	01
Digital glucometer	Measurement range- 60-100 mg/dl Sample volume – 2 microlit Result time- 10 sec Battery life- upto 600	01
Folin Wu tube	Capacity: 25 ml O.D. x Length (mm): 20 x 220 Material Glass (Borosilicate)	15
Hemostatic artery forceps	Length: 3.5" , 5 " Curvature: Straight Working Surface Style: Serrated Handle: Finger Ring Material: Stainless Steel	10
Cannula	Arterial and Venous cannula	20
Ampoule sealing machine	Hand operated for sealing cap. 1 ml to 10 ml (without Gas Cylinder)	01

Magnetic stirrer 500 ml	Voltage: 220 - 240 V. Power: 150 - 500 W. Operating Frequency: 50 Hz. Max Speed: 1200 rpm. Digital speed indicator	04
Magnetic stirrer 1000 ml	Voltage: 220 - 240 V. Power: 150 - 500 W. Operating Frequency: 50 Hz. Max Speed: 1200 rpm. Digital speed indicator	03

**Sd/-
Registrar
MAKAUT, WB**